

# Laser speckle imaging with an active noise reduction scheme

A.C. Völker<sup>1</sup>, P. Zakharov<sup>1</sup>, B. Weber<sup>2,3</sup>, F. Buck<sup>2</sup> and F. Scheffold<sup>1</sup>

<sup>1</sup>Department of Physics, University of Fribourg, 1700 Fribourg, Switzerland

<sup>2</sup>Division of Nuclear Medicine, University Hospital Zurich, Rämistrasse 100, 8091 Zurich, Switzerland

<sup>3</sup>Max Planck Institute for Biological Cybernetics, Spemannstraße 38, 72076 Tübingen, Germany  
[Andreas.Voelker@unifr.ch](mailto:Andreas.Voelker@unifr.ch), [Frank.Scheffold@unifr.ch](mailto:Frank.Scheffold@unifr.ch)

**Abstract:** We present an optical scheme to actively suppress statistical noise in Laser Speckle Imaging (LSI). This is achieved by illuminating the object surface through a diffuser. Slow rotation of the diffuser leads to statistically independent surface speckles on time scales that can be selected by the rotation speed. Active suppression of statistical noise is achieved by accumulating data over time. We present experimental data on speckle contrast and noise for a dynamically homogenous and a heterogeneous object made from Teflon. We show experimentally that for our scheme spatial and temporal averaging provide the same statistical weight to reduce the noise in LSI: The standard deviation of the speckle contrast value scales with the effective number  $N$  of independent speckle as  $1/\sqrt{N}$ .

**OCIS codes:** (110.6150) Speckle imaging; (170.7050) Turbid media; (170.3880) Medical and biological imaging; (110.4280) Noise in imaging systems.

## 1. Introduction

Laser speckle imaging (LSI) [1-6] is an optical technique that can be used to measure local dynamic properties in scattering media both at high temporal and spatial resolution. The underlying principle of LSI is equivalent to the one exploited in laser Doppler flowmetry (LDF), a method widely used in biomedical functional imaging of blood flow [2]. A moving object inside the tissue, such as a red blood cell, leads to a Doppler shift of the frequency of scattered light which is directly proportional to the velocity. Rather than analyzing Doppler frequency shifts LSI monitors temporal fluctuations of the scattered light in order to map dynamics. From a fundamental point of view both techniques are equivalent [1]. The frequency and time domain in quasi-elastic light scattering are connected by a simple Fourier transformation and thus completely interchangeable (Wiener-Khinchin theorem) [2, 3, 7, 8]. In other words if the spectrum of frequency shifts is known the distribution of temporal fluctuations can be calculated and vice versa. In LSI a digital camera records the surface intensity pattern of light reflected from a scattering medium. In the absence of movement (or for very short exposure time) there is no detectable fluctuation and the typical granular interference pattern, called speckle, has the same properties in the entire field of view. In the classical picture, summarized by Goodman in [9], the scattered field amplitudes are randomly distributed and consequently the intensity values follow a negative exponential distribution  $P(I)\langle I \rangle = \exp(-I/\langle I \rangle)$ . In an experiment the normalized standard deviation (or contrast) is:

$$K = \langle I \rangle^{-1} \left[ \beta \sum_{i=1}^N (I_i - \langle I \rangle)^2 / (N-1) \right]^{1/2}. \quad (1)$$

Here we have additionally introduced the so called coherence factor  $\beta$  that takes account of the finite size of the detection area and depolarization of light in the medium [10]; it approaches 1 only using a polarization filter and a point detector. For a fully developed static speckle  $K = \beta^{1/2}$ . Changes in the position of some objects lead to temporal intensity

fluctuations with a time constant given by the local velocity. Such time fluctuations are usually characterized by the intensity correlation function  $C(\tau) = \langle I(0)I(\tau) \rangle / \langle I \rangle^2 - 1$ . Instead of analyzing the temporal fluctuations in detail it is more convenient to measure a time averaged quantity simply by choosing a certain camera exposure time  $T$ . Then the average speckle contrast decreases according to  $K^2 = (\beta/T) \int_0^T 2(1-t/T)C(\tau)d\tau$ . For a simple exponential decay  $C(\tau) = \exp(-\tau/\tau_c)$  one finds [8, 11]:

$$K = \sqrt{\beta \frac{e^{-2x} - 1 + 2x}{2x^2}} \quad \text{with} \quad x = \frac{T}{\tau_c}. \quad (2)$$

The spatially resolved speckle contrast  $K$  is a sensitive probe of the dynamic properties. High contrast corresponds to slow movement while low contrast is found in areas of fast movement. In other words, areas with motion “blur” the speckle pattern, while static areas do not. In the so called Laser Speckle Contrast Analysis (LASCA) [12] approach an image is obtained from a simple analysis of the variance in a small area of typically  $N = 5 \times 5 = 25$  or  $N = 7 \times 7 = 49$  pixels. As a matter of fact calculation of the speckle contrast is limited to such small areas of a few pixels if a high spatial resolution is desired. We will call this area in the following “square size”. The choice of the square size is a compromise between statistical accuracy and spatial resolution. In such LASCA imaging the statistical noise is naturally very high, typically around  $1/\sqrt{N} = 15 - 20\%$  for a non-averaged single frame image.

## 2. Active noise reduction scheme

It would thus be of particular interest to find ways of improving the statistical accuracy without loss in spatial resolutions. A first step in this direction has been reported recently by Cheng et al. [5]. The authors calculate the speckle contrast from a series of intensity values recorded at the same position. Thus an image can be constructed without loss in spatial resolution. The fact that this method gives meaningful results is somewhat fortuitous as we will discuss below.

Here we propose a more rigorous approach to noise reduction based on an active speckle averaging scheme that ensures perfect ensemble averaging. The basic idea is to illuminate the sample with a random speckle field which can be varied externally using a rotating diffuser. Rather than using an expanded collimated beam the sample surface is illuminated with a dispersed laser beam that has passed a ground glass. The resulting divergent beam creates an effective speckle illumination on the sample surface placed at a distance of  $z = 20$  cm with a corresponding surface speckle size of  $\lambda z/\pi w \sim 25 \mu\text{m}$  for an incident beam size of approximately  $w = 2$  mm. Slowly rotating the ground glass leads to statistically independent realizations of the speckle pattern with characteristic temporal fluctuations on a time scale  $\tau_0$  which can be controlled externally by the motor speed. If  $\tau_0$  is chosen much larger than the desired exposure time  $T$  of the digital camera, subsequent speckle images taken at time intervals larger  $\tau_0$  are statistically independent.

The statistical properties of laser speckle created by a moving diffuser have been studied extensively in the past [9]. In contrast to our work, which exploits the speckle properties to construct a dynamic image, the aim of these previous studies was to develop efficient techniques to remove laser speckle in order to improve the static image quality. Nevertheless the general discussion of the speckle properties and contrast noise can be applied to both cases. Lowenthal and Joyeux, and later Yoshimura and Fujiwara, studied a combination of a motionless and a moving diffuser [14, 16]. Though similar, our experimental setup is somewhat more complex and we have thus not attempted to determine theoretically the speckle correlation time. Experimentally we find  $\tau_0 = 450\text{ms}$  for a rotation velocity of one rph.

### 3. Experiments and Results

We study the dynamic contrast resolution using a homogeneous block of solid Teflon and a home-made heterogeneous sample. This medical phantom mimics a liquid inclusion in solid tissue. It is obtained by milling a cylindrical hole of diameter  $D_0 = 3$  mm in a block of solid Teflon. A thin layer of 0.45 mm (accuracy 0.01 mm) separates the cylindrical inclusion from the interface that is imaged. The void is subsequently filled with a dispersion of 710 nm polystyrene particles in water. The particle concentration is adjusted to match the optical properties of the liquid to the solid (volume fraction ca. 1.3%). Thus our sample does not show any static scattering differences (Fig. 1).

A diode laser beam (785 nm, max. 50 mW) is incident on a ground glass (Edmund Optics), which is mounted on a motor. Light passing the ground glass is dispersed and illuminates the sample surface as an expanded light spot. The illuminated surface is then imaged with a camera objective (Schneider Kreuznach Xenoplan,  $f = 50$  mm, distance lens-image plane  $l = 80$  mm) onto the CCD chip of a digital camera (Pixelfly PCO, Germany). The camera resolution is 640x480 pixels (edge length of pixel  $r_p = 9.9$   $\mu\text{m}$ , half inch chip) and the full field of view is 16 by 12 mm. A crossed linear polarizer is put in front of the objective to eliminate spurious contributions from reflections and to increase the speckle contrast [13]. To correct for slightly inhomogeneous illumination we have divided the intensity value of each pixel by its mean (taken from 500 statistically independent measurements). In practice this only affects the results for square sizes 16x16 or larger, while for all others there is no notable difference with or without normalization.




Fig. 1. Experimental setup: A laser beam (785 nm) is incident on ground glass mounted on a motor (rotation velocity one rph). Light passing the ground glass is moderately divergent and illuminates the sample surface as an expanded light spot. The illuminated surface is imaged via a beam-splitter by the camera objective onto the CCD chip of the digital camera. Right: Heterogeneous sample obtained by milling a cylindrical void (diameter  $D_0 = 3$  mm) into a solid block of Teflon. The inclusion is filled with an aqueous suspension of 710 nm polystyrene latex spheres that matches the optical properties ( $l^* = 250 \pm 30$   $\mu\text{m}$ ) and creates dynamic contrast. A layer of thickness  $d = 0.45$  mm separates the inclusion from the imaging surface.

Due to scattering and dephasing in the ground glass the light spot on the surface displays a fully developed speckle pattern [14]. In the presence of the turbid media the characteristic surface speckle pattern is seen. This pattern is basically the same whether the beam is expanded by a lens or by the ground glass. As a matter of fact we expect no influence of the random diffuser on the surface speckle as long as the distance travelled by the light inside the sample is larger than the size of a surface speckle created by the diffuser. In our case this surface speckle size is approx. 25  $\mu\text{m}$  and the typical distance travelled by the light is the transport mean free path ( $l^* = 250 \pm 30$   $\mu\text{m}$  for Teflon and the colloidal suspension, as determined by diffusing wave spectroscopy [15]).

It is worthwhile to emphasize the marked difference to previous temporal-averaging schemes in LSI [5]. In our case the noise reduction is “active”. Every single image taken is statistically independent from the previous one. Passive time averaging schemes on the other

hand rely on experimental noise such as animal motion or other slow fluctuations that are beyond the control of the operator. Furthermore, the method proposed by Cheng et al. [5], is actually very similar to the empiric approach to average over a couple of images taken within a short time interval where the dynamics of the system (such as cerebral blood flow) can be assumed stationary [3, 6]. It is well known from digital camera based intensity correlation spectroscopy that in the limit of large numbers  $N$  the so called “divide and average scheme” provides the same results as the “average and divide” scheme [17, 18]. Time averaging in both cases however relies on the ergodicity of the medium (the possibility to obtain ensemble averages by time averaging) [19, 20]. In practice this means that consecutive pictures have to be statistically independent. This is only the case if the correlation function of each pixel  $C_{i,j}(\tau)$  has relaxed to zero between the acquisition of two images. Whether this is the case or not however depends on the camera frame rate and the unknown temporal behavior of  $C_{i,j}(\tau)$  which furthermore is different for each position (pixel  $i,j$ ).




Fig. 2. Characteristics of the laser speckle imaging setup obtained from a rigid Teflon block for different square sizes. a) mean speckle contrast  $K$  as a function of the speckle size (in units of the pixel size  $r_p$ ). b) Standard deviation of the contrast  $\sigma_K$  as a function of  $r_s$ . To achieve the same intensity for all apertures the exposure time was varied between 200 and 750 ms.

Some characteristics of our laser speckle imaging device are displayed in Fig. 2. A single image was taken for different imaging apertures from the homogeneous rigid Teflon block. The corresponding f-number ( $f/\#$ ) is the focal length  $f$  divided by the radius  $q$  of the aperture. The speckle size  $r_s$  on the CCD is then equal to  $r_s = 2l/k_0q = 4 \times f/\# \times l/k_0f$  [21]. Here  $l$  is the distance of the image to the lens and  $k_0$  is the wave number. The mean speckle contrast  $K$  and the contrast noise  $\sigma_K$  are shown for different box sizes. With increasing f-number the speckle size is increased and thus the mean contrast increases approaching the limit of a point detector  $r_s \gg r_p$ . At the same time the contrast noise increases since the number of speckle per area decreases (intensities on neighbouring pixels are more correlated). For all other experiments shown in this study we use  $f/\# = 16$  with a corresponding speckle size  $r_s = 12.8 \mu\text{m}$ , which is comparable to our CCD pixel size.

For the traditional LASCA scheme with 8x8 square size, Fig. 3(a), the contrast significantly decreases at the location of the liquid inclusion, overall however the signal remains very noisy. Fig. 3(b): A strikingly improved picture appears if we rotate the ground glass slowly with approximately one rph and construct an image from a series of measurements. 500 images were taken with 1 fps (frames per second). Thus the time interval between the single frames is larger than relaxation time due to motor rotation  $\tau_0$ . For a full resolution dynamic image (640x480 contrast values  $K$ ) the noise level can thus be estimated to  $1/\sqrt{N} = 4.5\%$  as compared to the 20% error in the 8x8 square size LASCA experiment at only 80x60 resolution.




Fig. 3. Three dimensional plot of a laser speckle contrast image of a liquid inclusion in a solid block of Teflon (scale inverted). a) LASCA image using 8x8 square size, image resolution 80x60 pixel averaged over 500 individual measurement. b) Same sample using the active noise reduction. c) Full resolutions 640x480 pixel with active noise reduction scheme.

We have performed a systematic study of the contrast noise for the case of a solid Teflon block. Since subsequent images are statistically independent, rotation introduces a “third dimension” in our averaging scheme. It should be of no importance whether the  $N$  pixels used to calculate the contrast  $K$  are taken along the  $x$ ,  $y$  or  $t$ -axes. Experimentally we indeed find that the noise in speckle contrast, characterized by its standard deviation  $\sigma_K = \sqrt{\sum_{i=1}^N (K_i - \langle K \rangle)^2 / (N-1)}$  scales with  $1/\sqrt{N}$  and  $N = x \cdot y \cdot \# \text{ frames}$  as long as  $N$  is sufficiently large ( $N > 100$ ), Fig. 4. For  $N > 10000$  the noise level saturates at a level well below 1%. We speculate that in this regime other sources of error become sizable such as spurious contributions of light not randomized by the diffuser, the inherent noise level of the CCD camera, fluctuations of the laser intensity or the beam point as well as contributions from ambient light or reflections at the optical elements.




Fig. 4. Left: Noise in speckle contrast, expressed by the standard deviation  $\sigma_K$ , as a function of the number of frames taken, different square sizes are displayed (where  $\# \text{ pixel}$  is the number of pixels in the corresponding square size). b)  $\sigma_K$  scaling as a function of “effective pixels”  $N$ . It is shown that spatial and temporal averaging give identical results.

#### 4. Summary and Conclusions

In summary we could demonstrate that our experimental scheme can efficiently suppress noise in LSI. Down to noise levels below 1% the accuracy that can be achieved is only limited by the total time of data acquisition. For a model experiment, as shown in the present examples, this time can be made arbitrarily long – leading to significant noise reduction as shown in Fig. 3(b),(c). Such improved accuracy in model experiments should turn out to be very useful in the theoretical and experimental discussion of general principles of light propagation and imaging in turbid media [22, 23]. In practical biomedical applications the number of images that can be acquired will be limited by the time available to record a

speckle image and the minimum exposure time set by the local dynamics. Furthermore the speckle correlation time imposed by the diffuser  $\tau_0$  has to be longer than the exposure time  $T$  (to avoid additional blurring of the speckle) but equal or shorter than the time between individual frames. In an actual biomedical imaging experiment this can be achieved with a reasonable exposure time of  $T=1$  ms [24] and selecting  $\tau_0 \sim 5$  ms. Capturing with 200 fps our approach can use 40 independent images within 0.2 seconds to construct a low-noise speckle image, a time interval sufficient to follow physiological changes in real time [6]. If combined with a moderate spatial average of  $4 \times 4$  pixels the noise level can be reduced to less than 4% over the full field of view. Note that the traditional LASCA spatial averaging scheme would require  $25 \times 25$  pixels to achieve the same statistical accuracy. We thus believe that our technically very simple scheme also offers significant advantages for practical implementations of LSI in biomedical imaging.

### Acknowledgements

Financial support by the Swiss National Science Foundation (Grant. No. 205321-104282/1) is gratefully acknowledged.

1. J. D. Briers, "Laser Doppler and time-varying speckle: A reconciliation," J. Opt. Soc. Am. A **13**, 345-350 (1996).
2. J. D. Briers, "Laser Doppler, speckle and related techniques for blood perfusion mapping and imaging," Physiological Measurement **22**, R35-R66 (2001).
3. A. K. Dunn, B. Hayrunnisa, M. A. Moskowitz, D. A. Boas, "Dynamic imaging of cerebral blood flow using laser speckle," J. Cereb. Blood Flow Metab. **21**, 195-201 (2001).
4. C. Ayata, Y. Ozdemir, A. Dunn, D. N. Atochin, P. L. Huang, V. R. Muzykantov, J. C. Murciano, D. A. Boas, M. A. Moskowitz, "Laser speckle-flowmetry: A novel two-dimensional technique for the study of cerebral blood flow in normal and ischemic mouse brain, in vivo," Stroke **34**, 251-251 (2003).
5. H. Cheng, e. al., "Modified laser speckle imaging method with improved spatial resolution," J. Biomed. Opt. **8**, 559-564 (2003).
6. B. Weber, C. Burger, M. T. Wyss, G. K. von Schulthess, F. Scheffold, A. Buck, "Optical imaging of the spatiotemporal dynamics of cerebral blood flow and oxidative metabolism in the rat barrel cortex," Eur. J. Neurosci. **20**, 2664-2670 (2004).
7. B. J. Berne, R. Pecora, *Dynamic Light Scattering* (John Wileys and Sons, New York, 1976).
8. R. Bandyopadhyay, A. S. Gittings, S. S. Suh, P. K. Dixon, D. J. Durian, "Speckle-visibility spectroscopy: A tool to study time-varying dynamics," Rev. Sci. Instrum. **76**, Art. No. 093110 (2005).
9. J. C. Dainty, *Laser speckle and related phenomena* (Springer-Verlag, Berlin ; New York, 1984).
10. L. F. Rojas, D. Lacoste, R. Lenke, P. Schurtenberger, F. Scheffold, "Depolarization of backscattered linearly polarized light," Journal of the Optical Society of America A-Optics Image Science and Vision **21**, 1799-1804 (2004).
11. E. Jakeman, E. R. Pike, "Intensity-Fluctuation Distribution of Gaussian Light," J. Phys. I **1**, 128-138 (1968).
12. J. D. Briers, S. Webster, "Laser Speckle Contrast Analysis (LASCA): A non-scanning, full-field technique for monitoring capillary blood flow," J. Biomed. Opt. **1**, 174-179 (1996).
13. F. C. Mackintosh, J. X. Zhu, D. J. Pine, D. A. Weitz, "Polarization Memory of Multiply Scattered-Light," Phys. Rev. B **40**, 9342-9345 (1989).
14. T. Yoshimura, K. Fujiwara, "Statistical properties of doubly scattered image speckle," J. Opt. Soc. Am. A **9**, 91-95 (1992).
15. D. J. Pine, D. A. Weitz, P. M. Chaikin, E. Herbolzheimer, "Diffusing-Wave Spectroscopy," Phys. Rev. Lett. **60**, 1134-1137 (1988).
16. S. Lowenthal, D. Joyeux, "Speckle Removal by a Slowly Moving Diffuser Associated with a Motionless Diffuser," J. Opt. Soc. Am. **61**, 847-851 (1971).
17. L. Cipelletti, D. A. Weitz, "Ultralow-angle dynamic light scattering with a charge coupled device camera based multispeckle, multitau correlator," Rev. Sci. Instrum. **70**, 3214-3221 (1999).
18. P. Zakharov, S. Bhat, P. Schurtenberger, F. Scheffold, "Multiple scattering suppression in dynamic light scattering based on a digital camera detection scheme," Appl. Opt. **to appear**, (2005).
19. P. N. Pusey, W. Vanmegen, "Dynamic Light-Scattering by Non-Ergodic Media," Physica A **157**, 705-741 (1989).
20. F. Scheffold, S. E. Skipetrov, S. Romer, P. Schurtenberger, "Diffusing-wave spectroscopy of nonergodic media," Phys. Rev. E **6306**, art. no.-061404 (2001).
21. T. Yoshimura, "Statistical properties of dynamic speckles," J. Opt. Soc. Am. A **3**, 1032-1054 (1986).
22. S. E. Skipetrov, I. V. Meglinskii, "Diffusing-wave spectroscopy in randomly inhomogeneous media with spatially localized scatterer flows," J. Exp. Theor. Phys. **86**, 661-665 (1998).
23. A. C. Völker, P. Zakharov, B. Weber, A. Buck, F. Scheffold, "Dynamic contrast resolution in laser speckle imaging," in preparation (2005).
24. S. Yuan, A. Devor, D. A. Boas, A. K. Dunn, "Determination of optimal exposure time for imaging of blood flow changes with laser speckle contrast imaging," Appl. Opt. **44**, 1823-1830 (2005).